

# POTTY-CHECK® IN-SERVICE REVIEW

## PROPER INSTALLATION OF THE POTTY-CHECK SYSTEM WHEN USED WITH A BASIC-CHECK® CONTROL UNIT

For more information, consult Basic-Check System User Guide



**A.** Remove Sensormat® from plastic container. Use the adhesive backing on the Sensormat to properly position it. It is usually placed between the two seat bumpers, but may be placed on the top edge of the toilet bowl. Locate it forward for lighter patients and rearward for heavier patients.



**B.** Plug the Potty-Check Sensormat's phone-style jack into the Basic-Check control unit.

**IMPORTANT:** Use only Sensormats manufactured by Bed-Check® with the Basic-Check control unit.



**C.** Using either the integral hanger or Velcro® backing, hang the Basic-Check on a bathroom handrail or wall where readily accessible to the caregiver, and out of reach of the patient. Record date Sensormat put in use on cord tag.



**D.** Lower toilet seat and help patient onto toilet. The control unit will arm itself automatically when the patient's weight is sensed. The alarm will beep one time, indicating that monitoring has begun.



**E.** If the patient exits the toilet seat, the alarm will sound immediately. Alarm requires immediate intervention.



**F.** When an alarm has sounded, first assist the patient. The alarm may be silenced by seating the patient again, automatically putting Basic-Check into monitor mode. Otherwise silence the alarm by pushing the Reset button, putting Basic-Check into stand-by mode.



**G.** To remove the patient from the toilet seat, allow the Basic-Check to alarm and press the Reset button on the Basic-Check one time. This puts the Basic-Check into stand-by mode.

### IMPORTANT:

Check for System Effectiveness Daily and/or at shift change.

Always check correct positioning of Sensormat after any alarm.

Use Bed-Check products only to maintain Bed-Check warranty and to ensure system compatibility.

The physical characteristics, habits and factors that make each patient at risk for falls are different and may change daily. Please reassess as necessary.

**STANLEY**



Reducing the Risk of Falls Through  
Innovative Technology Since 1978

P.O. Box 80238, Lincoln, NE 68501  
1-800-523-7956 • [www.bedcheck.com](http://www.bedcheck.com)