

# POTTY-CHECK® IN-SERVICE REVIEW

## PROPER INSTALLATION OF THE POTTY-CHECK SYSTEM WHEN USED WITH A CHAIR-CHECK® II CONTROL UNIT

For more information, consult Chair-Check II System User Guide



**A.** Remove Sensormat® from plastic container. Use the adhesive backing on the Sensormat to properly position it. It is usually placed between the two seat bumpers, but may be placed on the top edge of the toilet bowl. Locate it forward for lighter patients and rearward for heavier patients.



**B.** Plug the Potty-Check Sensormat's phone-style jack into the Chair-Check II control unit. Delay should be set at 0 or 1 second (see User Guide).

**IMPORTANT:** Use only Sensormats manufactured by Bed-Check® with the Chair-Check II control unit.



**C.** Using a Chair-Check wire holster, attach the Chair-Check II to a bathroom handrail where readily accessible to the caregiver, and out of reach of the patient.

Record date Sensormat put in use on cord tag.



**D.** Lower toilet seat and help patient onto toilet. The control unit will arm itself automatically when the patient's weight is sensed. The alarm will beep one time, followed by a blinking red light located next to the Reset button.



**E.** If the patient/resident exits the toilet seat, the alarm will sound. Alarm requires immediate intervention.



**F.** After the alarm has sounded, first assist the patient and then press the Reset button. This silences the alarm and puts the Chair-Check II into stand-by mode.



**G.** To remove the patient from the toilet seat, press Reset on the Chair-Check II one time. This puts the unit into stand-by mode.



**H.** If the patient/resident has not been removed from the seat in one minute after Reset is pressed, Chair-Check II will resume monitoring, indicated by one beep and the blinking red light.

### IMPORTANT:

Check for System Effectiveness Daily and/or at shift change.

Always check correct positioning of Sensormat after any alarm.

Use Bed-Check products to maintain Bed-Check warranty and to ensure system compatibility.

The physical characteristics, habits and factors that make each patient at risk for falls are different and may change daily. Please reassess as necessary.

**STANLEY**



P.O. Box 80238, Lincoln, NE 68501  
1-800-523-7956 • [www.bedcheck.com](http://www.bedcheck.com)

©2009 Stanley Security Solutions, Inc. DOC-96-70010-AA