

The Model II-1F/AC, Model V, Model Vr, Chair-Check and Chair-Check II Control Units are designed and manufactured to be used with the Bed-Check Sensormats. Danger: This monitor may not be compatible with mats manufactured by companies other than Bed-Check Corp. Use of incompatible mats may result in patient injury or death, and work thd product warranty

Please call Bed-Check immediately if you experience problems.



Reducing Falls Through
Innovative Technology Since 1978

For the location and phone number of your area Bed-Check distributor or for technical assistance, contact Bed-Check Corporation direct at:

1-800-523-7956

FAX: 1-918-582-9828
Bed-Check Corporation
P.O. Box 170
Tulsa, Oklahoma 74101-0170
www.bedcheck.com